



Standard Practice for Determination of a Pooled Limit of Quantitation¹

This standard is issued under the fixed designation D6259; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers the determination of a lower quantitative limit for a test method for an analyte. The determined lower limit is hereinafter referred to as the *pooled limit of quantitation*.

1.2 Applicable test methods will produce test results greater than zero. Examples are those test methods that measure sample composition.

2. Referenced Documents

2.1 *ASTM Standards*:²

[E456 Terminology Relating to Quality and Statistics](#)

[E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method](#)

2.2 *ASTM Adjuncts*:

D2PP, Determination of Precision and Bias Data for Use in Test Methods for Petroleum Products³

3. Terminology

3.1 *Definitions*:

3.1.1 *pooled limit of quantitation, n*—level of property or concentration of analyte above which quantitative test results can be obtained with a specified degree of confidence. See [3.2.1](#) for acronym.

3.1.2 *repeatability conditions, n*—conditions under which test results are obtained with the same test method in the same laboratory by the same operator with the same equipment in the shortest practical period of time using test units or test specimens taken at random from a single quantity of material that is as nearly homogeneous as possible (see [10.3](#) of Practice [E691](#).)

NOTE 1—The *same operator, same equipment* requirement means that for a particular step in the measurement process the same combination of operator and equipment is used for every test result. Thus, one operator

¹ This practice is under the jurisdiction of Committee D02 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.94 on Coordinating Subcommittee on Quality Assurance and Statistics.

Current edition approved May 1, 2004. Published June 2004. Originally approved in 1998. Last previous edition approved in 1998 as D5259-98. DOI: 10.1520/D6259-98R04.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Withdrawn and no longer available from ASTM International.

may prepare the test specimens, a second measure the dimensions, and a third measure the mass in a test method for measuring density.

NOTE 2—By *in the shortest practical period of time* is meant that the test results, at least for one material, are obtained in a time period not less than in normal testing and not so long as to permit significant change in test material, equipment, or environment. See Terminology [E456](#).

3.2 *Acronyms*:

3.2.1 *PLOQ, n*—pooled limit of quantitation.

4. Summary of Practice

4.1 Determine the standard deviation of a test result, under repeatability conditions, at progressively higher levels of the analyte until the ratio of measured level to standard deviation becomes greater than ten and remains so.

5. Significance and Use

5.1 In a single laboratory, the limit of quantitation, LOQ, equal to ten standard deviations has been recommended.⁴ A test result at this LOQ has an uncertainty of $\pm 30\%$ at the 99% confidence level. Similarly here, as a general estimate, the PLOQ, equal to ten repeatability standard deviations is recommended. A test result at this PLOQ has an uncertainty of $\pm 30\%$ at the 99% confidence level.

5.2 Values below the PLOQ are deemed to be too uncertain for meaningful use in commerce, or in regulatory activities.

5.3 Many test methods never find application outside their PLOQ. However, in the quest for ever more sensitive procedures, it can become difficult to distinguish an analytical response from background noise with the technology at hand. Test methods defective in design or poorly executed may also function outside their PLOQ.

6. Procedure

6.1 Make the preparations outlined in [6.2](#), then carry out one of the procedures described in [6.3](#).

6.2 *Preparations*:

6.2.1 *Select Test Levels*—Decide the objective of the test method, the range of typical samples it is expected to cover. Name a set of test levels covering this range and spaced to bracket the PLOQ. In some cases, the PLOQ will be well below the useful range. Then, it is only necessary to determine a *less than* value.

⁴ Keith, L. H., et al, "Principles of Environmental Analysis," *Analytical Chemistry*, American Chemical Society, Vol 55, 1983, p. 227.

TABLE 1 Interlaboratory Study Examples

Sample	Mean	Within Laboratories		Ratio Mean/Standard Deviation
		Standard Deviation	Degrees of Freedom	
S8 ^A	110.0	44.72	10	2.5
S1 ^A	640.0	100.0	10	6.4
S3 ^A	870.0	89.44	10	9.7
S6	1180	94.87	10	12.4
S2	1272	78.17	9	16.3
S7	2505	136.0	10	18.4
S4	3165	125.0	8	25.3
S5	3338	112.4	10	29.6

^A Ratio of mean to standard deviation is less than ten and is, therefore, lower than the PLOQ.

TABLE 2 Abbreviated Example—One Laboratory

Run/Sample	A	Blank	A-Blank	B	Blank	B-Blank
1	800	10	790	1050	20	1030
2	900	5	895	1250	20	1230
3	900	10	890	1250	5	1245
4	800	0	800	1300	0	1300
5	750	0	750	1200	5	1195
6	950	20	930	1100	10	1090
7	1000	5	995	1150	5	1145
Mean			864.3			1176.4
Standard Deviation			87.29			94.06
Ratio Mean/Standard Deviation	-		9.9 ^A	-		12.5

^A Ratio of the mean of sample A to its standard deviation is less than ten and is, therefore, lower than the LOQ.

NOTE 3—Mean and standard deviation data from experienced laboratories, archived research reports, and known limitations of the test method or equipment can give a preliminary notion of the PLOQ. An LOQ can be estimated in a single laboratory.

6.2.2 *Select Sample Materials*—Normally, use sample materials that are typical of those to which the test method is applied. In special cases, the method of standard additions (spiking) can be necessary to achieve the selected test levels. Synthetic blends may be required because of cost or other practical considerations.

6.2.2.1 Plan to determine the PLOQ of each important test result of a multicomponent test. The PLOQs are not additive.

6.2.3 *Select Number of Laboratories*—Determine the PLOQ with the precision during an interlaboratory study so that it will tend to be representative of industry wide experience.

NOTE 4—Just as it is possible to obtain a preliminary estimate of test method repeatability by performing a study in one laboratory, so it is possible to estimate the LOQ of the test method in one laboratory.

6.2.4 *Determine Number of Runs Per Sample*—If a blank measurement is required to calculate the measured level of analyte, the analyst must obtain a separate blank measurement for each sample aliquot analyzed.

NOTE 5—At least six degrees of freedom are required. When a precision study is done, the relevant standard deviation is the repeatability standard deviation. The number of degrees of freedom here equals the number of participating laboratories times the number of repetitions less one by each laboratory. When one laboratory is used, the relevant standard deviation is that obtained by repeating the test seven or more times on portions of one sample.

6.3 Continue with 6.3.1. However, since the PLOQ depends on the attainable precision, there can be cases where a laboratory-specific precision and LOQ are needed by an individual laboratory for a special purpose. If the decision is made to determine the LOQ in a single laboratory, proceed to 6.3.2.

6.3.1 *Interlaboratory*—Conduct the interlaboratory study to determine precision in a normal fashion making sure that appropriate samples are included. Proceed to the calculation steps (Section 7).

6.3.2 *One Laboratory*:

6.3.2.1 Select an initial sample. This initial sample and a second sample (6.2.3) should be selected to bracket an expected LOQ broadly.

6.3.2.2 Run the entire test method on seven aliquots and calculate mean, standard deviation, and the ratio mean/standard deviation.

6.3.2.3 Select a second sample and repeat the determinations and calculations.

6.3.2.4 Repeat 6.3.2.1-6.3.2.3 and, for example by a binary search process, converge on the LOQ. Exceptional situations include those where cases of practical interest are either all above or all below the LOQ as evidenced by the mean to standard deviation ratios.

6.4 An LOQ should be redetermined when there are substantive changes in the operator experience level, test method, instrumentation, or the sample types to which it is deemed applicable, and another interlaboratory study becomes necessary.

7. Calculation

7.1 Arrange the means and corresponding standard deviations in ascending order of the mean.

7.2 Divide each mean by the corresponding standard deviation to obtain a ratio.

7.3 The level at which these ratios becomes greater than ten and remains so is the PLOQ. A more precise value may be obtained by regression but not by extrapolation.

8. Reporting Low-level Data

8.1 Should a test result be reported that is below the PLOQ, give the PLOQ in parentheses after the data, for example, chlorine content = 110 mg/kg (PLOQ=1000 mg/kg).

8.2 A statement about the PLOQ should appear in the test method as a footnote to the Scope.

9. Examples

9.1 Examples from an interlaboratory study are provided in Table 1.

9.1.1 The PLOQ itself might be estimated at 1000 mg/kg. Tables similar to Table 1 are generated by the ADJ6300 computer software program.

9.2 *Study in One Laboratory*—The abbreviated example in Table 2 illustrates one possible way to tabulate the data.

9.2.1 It would seem that the LOQ is at or above the level of Sample A. Nevertheless, it is recommended that at least another sample at a lower level be run to make sure this is not an outlier.

10. Keywords

10.1 limit of quantitation; pooled limit of quantitation

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).